

SP046

## NMR of Glycolipids

### Materials

Methanol- $d_4$   
Choloroform- $d$   
N<sub>2</sub> bath [SOP SP031]  
NMR tubes  
13 x 100 mm tubes + caps

### Protocol

1. \_\_\_\_ Obtain 1-5 mg of dry sample in an acetone-rinsed 13 x 100 mm tube. Do the same for a control sample. (Note 1)
2. \_\_\_\_ Resuspend each sample in 1 ml CDCl<sub>3</sub>/CD<sub>3</sub>OD (2:1), and dry down.
3. \_\_\_\_ Repeat step (2).
4. \_\_\_\_ Resuspend each sample in 1 ml deuterated solvents again, and transfer to NMR tubes. Parafilm tubes and keep at 4° C or -20° C before use. (Note 2)
5. \_\_\_\_ In the basement of the Chemistry building, clean each tube with a kimwipe and EtOH, and ensure tube is straight by inserting into the glass gauge.
6. \_\_\_\_ For JS-300 work, log in user name *jdw* and password *nmr4CSU*. In the command window enter **e**. Make certain the air is on, insert the spinner with the sample, and enter **i**.
7. \_\_\_\_ Click on **<setup>**, select **<CDCl<sub>3</sub>>** as the solvent system, and enter **su** to initialize the hardware. (Note 3)
8. \_\_\_\_ To load the shim program enter **rts**, **shimfile**, then **setallshims**.
9. \_\_\_\_ Select **<acqui>** and click on **<lock>**. Ensure that the spin is at 20 Hz. Click **<lock off>** before adjusting any parameters. Maximize the *lock gain*. (Note 4)
10. \_\_\_\_ Begin obtaining a step function by increasing  $z_0$  by 1 from about 300 until you see a plateau "step". (Note 5)
11. \_\_\_\_ Click **<lock on>** once a good step function is obtained.
12. \_\_\_\_ Select **<shim>** to fine tune the signal. Adjust  $z_1$  and  $z_2$  alternatively, reducing the *lock gain* if the signal reads above 100%. (Note 6)
13. \_\_\_\_ Enter **nt=128** or greater (a multiple of 4) as the number of scans, then **ga** to begin acquisition. (Note 7)
14. \_\_\_\_ Once the spectrum is finalized, save the data with the **svf** command followed by the file name at the next prompt, making sure to exclude spaces.

15. \_\_\_\_ To begin processing the spectrum, enter **aph** (phase correction), and **vsadj** to adjust peak height. Ensure **<Display>** and **<Interactive>** have been selected to enable further manipulation.
16. \_\_\_\_ Select **<Full>** and **<dscale>**. (Note 8)
17. \_\_\_\_ Expand a region of interest by setting the red borders with the left and right mouse buttons, then select **<Expand>**.
18. \_\_\_\_ Enter **atext** ('experiment description') for a general description of the plot. Type **atext(file)** to add the file name to the printout.
19. \_\_\_\_ Select **<th>** to set the level above which peaks will be labeled.
20. \_\_\_\_ Adjust peak height by clicking on the spectrum line with the middle button and slowly moving upward.
21. \_\_\_\_ To print, select **<main menu>**, **<display>**, and **<plot>**, then **<plot>**, **<scale>**, **<all parameters>**, **<peaks>**, and **<page>**.
22. \_\_\_\_ Enter **e**, remove sample, then enter **i**. Within the desktop click and hold the right mouse button, slide down to *log out*, and release.
23. \_\_\_\_ Transfer samples to 13 x 100 mm tubes and dry down.
24. \_\_\_\_ Resuspend each in 1 ml normal, non-deuterated 2:1 solvent, and dry down.
25. \_\_\_\_ Repeat step (24) once more. Samples should be stored at 4° C until further QC.
26. \_\_\_\_ Resuspend in 1 ml 2:1 and transfer back to bulk sample for weight analysis.

## Notes

- (1) The Inova-400 can detect down to about 0.5 mg sample, but at least 3.0 mg will be needed for the JS-300 NMR.
- (2) An NMR tube rack can be used, or the tubes may be held and transported in an Erlenmeyer flask stuffed with kimwipes to prevent jostling.
- (3) There is not a selection for the dual solvent used for TDM or SL, but this is the closest.
- (4) It may be more efficient to use parameters from the previous run, in which case the *lock gain* can be set to that value now. Leave *lock phase* alone. *Lock power* should not exceed 36 as excessive power will make it difficult to shim.
- (5) Standard TDM will optimize between 300 and 350. Double check optimization by moving above and below this parameter by increments of 1. The best  $z_0$  will be that with the highest variation between the upper and lower plateau, which also seems to remain constant. Record this number before moving on to shim.
- (6) A signal which fluctuates between 60% and 100% is good, and between 80% and 100% is better. Continue to adjust *lock gain* down by 1 until the course signal will not exceed 100%, then click close. Record these parameters for future reference.

- (7) The number of scans can be set to any multiple of the batch scan (BS) number. Monitor the spectrum as you go by entering **wft**.
- (8) The tallest peak on the right is the tetramethyl silane (TMS) standard. Before clicking on **<dscale>**, expand that region, right click in center of peak, and click **<ref>**, then enter '0.' Once you click on **<dscale>** it will reset this peak as the zero reference.